UNITED STATES DEPARTMENT OF COMMERCE United States Patent and Trademark Office Address: COMMISSIONER FOR PATENTS P.O. Box 1450 Alexandria, Virginia 22313-1450 www.uspto.gov

APPLICATION NO.	FILING DATE	FIRST NAMED INVENTOR	ATTORNEY DOCKET NO.	CONFIRMATION NO.
10/560,722	12/15/2005	Wojtek Sudol	US030151US	8533
24737 7590 12/24/2008 PHILIPS INTELLECTUAL PROPERTY & STANDARDS P.O. BOX 3001			EXAMINER	
			CATTUNGAL, SANJAY	
BKIARCLIFF I	BRIARCLIFF MANOR, NY 10510		ART UNIT	PAPER NUMBER
			3768	
			MAIL DATE	DELIVERY MODE
			12/24/2008	PAPER

Please find below and/or attached an Office communication concerning this application or proceeding.

The time period for reply, if any, is set in the attached communication.

	Application No.	Applicant(s)			
Office Action Comments	10/560,722	SUDOL, WOJTEK			
Office Action Summary	Examiner	Art Unit			
	SANJAY CATTUNGAL	3768			
The MAILING DATE of this communication app Period for Reply	ears on the cover sheet with the c	orrespondence address			
A SHORTENED STATUTORY PERIOD FOR REPLY IS SET TO EXPIRE 3 MONTH(S) OR THIRTY (30) DAYS, WHICHEVER IS LONGER, FROM THE MAILING DATE OF THIS COMMUNICATION. - Extensions of time may be available under the provisions of 37 CFR 1.136(a). In no event, however, may a reply be timely filed after SIX (6) MONTHS from the mailing date of this communication. - If NO period for reply is specified above, the maximum statutory period will apply and will expire SIX (6) MONTHS from the mailing date of this communication. - Failure to reply within the set or extended period for reply will, by statute, cause the application to become ABANDONED (35 U.S.C. § 133). Any reply received by the Office later than three months after the mailing date of this communication, even if timely filed, may reduce any earned patent term adjustment. See 37 CFR 1.704(b).					
Status					
1)⊠ Responsive to communication(s) filed on <u>19 Se</u>	eptember 2008.				
<i>i</i> —		secution as to the merits is			
	Since this application is in condition for allowance except for formal matters, prosecution as to the merits is closed in accordance with the practice under <i>Ex parte Quayle</i> , 1935 C.D. 11, 453 O.G. 213.				
ologod in addordance with the practice and c	x parto Quayro, 1000 0. 5 . 11, 10	0.0.210.			
Disposition of Claims					
4)⊠ Claim(s) <u>1-29</u> is/are pending in the application.					
4a) Of the above claim(s) is/are withdrawn from consideration.					
5) Claim(s) is/are allowed.					
6)⊠ Claim(s) <u>1-29</u> is/are rejected.					
7) Claim(s) is/are objected to.					
· ·	election requirement				
8) Claim(s) are subject to restriction and/or election requirement.					
Application Papers					
9) The specification is objected to by the Examiner.					
10)⊠ The drawing(s) filed on <u>15 December 2005</u> is/are: a)⊠ accepted or b)□ objected to by the Examiner.					
Applicant may not request that any objection to the drawing(s) be held in abeyance. See 37 CFR 1.85(a).					
Replacement drawing sheet(s) including the correction is required if the drawing(s) is objected to. See 37 CFR 1.121(d).					
11) The oath or declaration is objected to by the Examiner. Note the attached Office Action or form PTO-152.					
Priority under 35 U.S.C. § 119					
 12) Acknowledgment is made of a claim for foreign priority under 35 U.S.C. § 119(a)-(d) or (f). a) All b) Some * c) None of: 1. Certified copies of the priority documents have been received. 					
Certified copies of the priority documents	have been received in Application	on No			
Copies of the certified copies of the prior	3. Copies of the certified copies of the priority documents have been received in this National Stage				
application from the International Bureau	application from the International Bureau (PCT Rule 17.2(a)).				
* See the attached detailed Office action for a list of the certified copies not received.					
Attachmont/e\					
Attachment(s) 1) X Notice of References Cited (PTO-892) 4) Interview Summary (PTO-413)					
2) Notice of Draftsperson's Patent Drawing Review (PTO-948) Paper No(s)/Mail Date					
3) Information Disclosure Statement(s) (PTO/SB/08) 5) Notice of Informal Patent Application					
Paper No(s)/Mail Date 6) Other:					

DETAILED ACTION

Response to Arguments

Applicant's arguments with respect to claims 1-29 have been considered but are moot in view of the new ground(s) of rejection.

Claim Rejections - 35 USC § 102

The following is a quotation of the appropriate paragraphs of 35 U.S.C. 102 that form the basis for the rejections under this section made in this Office action:

A person shall be entitled to a patent unless -

(b) the invention was patented or described in a printed publication in this or a foreign country or in public use or on sale in this country, more than one year prior to the date of application for patent in the United States.

Claims 1-3, 8-11, 16-18, and 23-25 are rejected under 35 U.S.C. 102(b) as being anticipated by U. S. Patent No. 6,381,197 to Savord et al.

Regarding Claims 1-3, 8-11, 16-18, and 23-25, Savord teaches an ultrasonic imaging apparatus comprising: an ultrasonic probe; a plurality of acoustic elements configured and arranged to form a two-dimensional array, said two-dimensional array configured and adapted to fit within the ultrasonic probe wherein said plurality of acoustic elements includes at least one active acoustic element capable of generating an acoustic pulse and/or receiving an echo signal (Abstract, Claim 1 and 13); and a beam controller coupled to said two-dimensional array, said beam controller capable of driving the at least one active acoustic element to produce said acoustic pulse for impinging an acoustic target to generate at least one echo signal, and having

associated circuitry capable of controlling directional movement of said acoustic pulse by tractor treading. (Abstract, Claim 1 and 13; and Col. 2 lines 25-65)

Claim Rejections - 35 USC § 103

The following is a quotation of 35 U.S.C. 103(a) which forms the basis for all obviousness rejections set forth in this Office action:

(a) A patent may not be obtained though the invention is not identically disclosed or described as set forth in section 102 of this title, if the differences between the subject matter sought to be patented and the prior art are such that the subject matter as a whole would have been obvious at the time the invention was made to a person having ordinary skill in the art to which said subject matter pertains. Patentability shall not be negatived by the manner in which the invention was made.

Claims 4-7, 12-15, 19-22, and 26-29 are rejected under 35 U.S.C. 103(a) as being unpatentable over Savord in view of U. S. Patent No. 5,938,612 to Kline-Schroder et al. (Kline)

Regarding Claims 5, 13, 20, and 27 Savord teaches all of the above claimed limitations but does not expressly teach the use of cylindrical transducer array configuration.

Kline teaches the use of cylindrical transducer array configuration.

It would have been obvious to one of ordinary skill in the art at the time of invention to modify Savord with a cylindrical transducer array configuration as taught by Kline since such a setup would result in a transducer array that could be inserted into blood vessels and inter-cavity tubes. (Col. 11 lines 59-65)

Regarding Claims 4, 6, 7, 12, 14, 15, 19, 21, 22, 26, 28, and 29 Kline teaches the use of planar and convex transducer array configuration (Figs. 11-14).

Conclusion

Applicant's amendment necessitated the new ground(s) of rejection presented in this Office action. Accordingly, **THIS ACTION IS MADE FINAL**. See MPEP § 706.07(a). Applicant is reminded of the extension of time policy as set forth in 37 CFR 1.136(a).

A shortened statutory period for reply to this final action is set to expire THREE MONTHS from the mailing date of this action. In the event a first reply is filed within TWO MONTHS of the mailing date of this final action and the advisory action is not mailed until after the end of the THREE-MONTH shortened statutory period, then the shortened statutory period will expire on the date the advisory action is mailed, and any extension fee pursuant to 37 CFR 1.136(a) will be calculated from the mailing date of the advisory action. In no event, however, will the statutory period for reply expire later than SIX MONTHS from the date of this final action.

Any inquiry concerning this communication or earlier communications from the examiner should be directed to SANJAY CATTUNGAL whose telephone number is (571)272-1306. The examiner can normally be reached on 9:30 - 5:00 pm.

If attempts to reach the examiner by telephone are unsuccessful, the examiner's supervisor, Long Le can be reached on (571)272-0823. The fax phone number for the organization where this application or proceeding is assigned is 571-273-8300.

Application/Control Number: 10/560,722 Page 5

Art Unit: 3768

Information regarding the status of an application may be obtained from the Patent Application Information Retrieval (PAIR) system. Status information for published applications may be obtained from either Private PAIR or Public PAIR. Status information for unpublished applications is available through Private PAIR only. For more information about the PAIR system, see http://pair-direct.uspto.gov. Should you have questions on access to the Private PAIR system, contact the Electronic Business Center (EBC) at 866-217-9197 (toll-free). If you would like assistance from a USPTO Customer Service Representative or access to the automated information system, call 800-786-9199 (IN USA OR CANADA) or 571-272-1000.

SPC

/Long V Le/ Supervisory Patent Examiner, Art Unit 3768